

English EC Declaration of Conformity

We, Gram Commercial declare under sole responsibility that the following products:

Range: Model: Rerigerant: Product description: Valid from (Year/Week): BioUltra UL570 HC: R290 & R170 (R601 as additive) HFC: R404A & R508B (R601 as additive) Extreme low temperature freezers for storing biomedical materials 2022/02

To which this declaration relates, is in compliance with all the applicable essential requirements, and other provisions of the European Council Directive and regulations.

Directives and Regulations of the European Parliament and of the Council:

Machinery Directive 2006/42/EC

- ATEX Directive 2014/34/EU

- Pressure Equipment Directive 2014/68/EU

- Low Voltage Directive 2014/35/EU

- EMC Directive 2014/30/EU

- RoHS Directive 2011/65/EU

- REACH EC No.1907/2006

- F-Gas Regulation (EU) No 517/2014

Product compliance has been demonstrated on the basis of:

| Harmonized Standards: | Text: |
|--|---|
| EN 60601-1:2006 | Medical electrical equipment – Part 1: General requirements for basic safety and essential performance |
| EN 60601-1-2:2015 | Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests |
| EN 61010-1:2010 | Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements |
| EN 61326-1:2013 | Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements |
| EN IEC 61000-3-2:2019 | Electromagnetic compatibility (EMC) – Part 3-2: Limits – Limits for harmonic current emissions (equipment input current =16 Å per phase) |
| EN 61000-3-3:2013 | Electromagnetic compatibility (EMC) – Part 3-3: Limits – Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current <= 16 A per phase and not subject to conditional connection |
| EN IEC 60079-0:2018 EN IEC 60079-0:2018/AC:2020 | Explosive atmospheres – Part 0: Equipment – General requirements |
| EN 60079-7:2015 EN 60079-7:2015/A1:2018 | Explosive atmospheres – Part 7: Equipment protection by increased safety "e" |
| EN 60079-11:2012 | Explosive atmospheres – Part 11: Equipment protection by intrinsic safety "i" |
| EN IEC 60079-15:2019 | Explosive atmospheres – Part 15: Equipment protection by type of protection "n" |
| EN 60079-18:2015 | Explosive atmospheres – Part 18: Equipment protection by encapsulation "m" |
| EN ISO 3744:2010 | Acoustics – Determination of sound power levels and sound energy levels of noise sources using sound pressure – Engineering methods for an essentially free field over a reflecting plane |
| EN ISO 9001:2015 | Quality management systems – Requirements |
| EN ISO 14001:2015 | Environmental management systems – Requirements with guidance for use |

Gram Commercial Aage Grams Vej 1 DK-6500 Vojens Telephone: + 45 73 20 13 00

Vojens, 13.1.2022

John B. S. Petersen Approval Manager